

**STREAM FOUNDATION**  
**Tender Ref no STREAM/EQ/01/010/2025**  
**CORRIGENDUM-6**

Sl no	Tender Clause	Original clause	Revised clause
1	Annexure -B Point 4	The system should allow for measurements in inert, oxidizing, reducing, vacuum, atmospheres.	The system should allow for measurements in inert, oxidizing, <b>reducing, atmospheres.</b>
2	Annexure –B point 4	Automatic evacuation and gas filling between measurements should be available.	<b>Point cancelled</b>
3	Annexure –B point 6	A set of standard samples (sapphire discs) for Cp calibration should be available for alumina and platinum crucibles.	A set of standard samples (sapphire discs) for Cp calibration should be available for <b><u>alumina crucibles.</u></b>