

Corrigendum 3
Tender Ref no STREAM/EQ/011/01/2025
TGA/DTA machine

Sl no.	Tender clause-Annexure -B	Original Clause	Revised Clause
1	2	Temperature resolution 0.001 °C	Resolution of 0.01°C minimum to be supplied across the entire range of temperature
2	4	Thermocouple : Type S	Thermocouple to be Pt/Pt-Rh or equivalent
3	4	The baseline drift must be less than 2 µV	“Bidder to specify baseline drift when quoting. High drifts can be cause for rejection during technical evaluation”
4	8	The equipment and the accessories / standard specimens should be shall be provided that are traceable to NIST.	The equipment and the accessories / standard specimens should be shall be provided that are traceable to NIST minimum of 2.

5	2	Temperature accuracy 0.2 °C	Temperature accuracy 1 °C or better
6	5	Software should be able to export raw data as EXCEL/CSV file, TMDSC, Cp	Software should be able to export raw data as EXCEL/CSV file.