

Corrigendum-3
Tender Ref no STREAM/EQ/012/01/2025
TMA machine

Sl no.	Tender clause-Annexure -B	Original Clause	Revised/added clause
1	1	Added Point	“Bidders can optionally quote 2 standalone machines to cover the temperature range”
2	2	Temperature accuracy: $\pm 1\text{ }^{\circ}\text{C}$ Temperature precision: $0.3\text{ }^{\circ}\text{C}$ Temperature resolution: $0.001\text{ }^{\circ}\text{C}$	Temperature accuracy: $\pm 2\text{ }^{\circ}\text{C}$ Temperature precision: $1\text{ }^{\circ}\text{C}$ Temperature resolution: $0.01\text{ }^{\circ}\text{C}$ or better
3	3	Measuring range: $\pm 500\text{ }\mu\text{m}$	Measuring range: $\pm 2500\text{ }\mu\text{m}$
4	3	Δl Resolution : 0.125 nm	Delta L resolution = 0.125nm or better.
5	3	The drift of the force signal should be smaller than 0.2 mN	The noise of the force measurement should be smaller than 0.5 mN (peak to peak)
6	3	Added point	“Stress and Strain control required

7	3	Added point	“Forced air or cooling system can be provided for faster cooling”
8	3	Heating rate: 0.01 to 100 °C/min	Heating rate: 0.01 to 50 °C/min