

**Prebid online meeting questions and reply**  
**Tender Ref no STREAM/EQ/013/01/2025**  
**Brinell Hardness Testing Machine**

<b>Sl no.</b>	<b>Tender clause-Annexure -B</b>	<b>Question</b>	<b>Reply</b>
1	1	For automatic measurements would you require computer to be provided?	Automatic measurements can be provided using a computer or display
2	2	What are the loads required at 2.5mm ball dimensions	Refer corrigendum 3
3	2	Is CRM required for 2.5mm ball	2.5mm ball removed hence CRM will not be required
4	7	Clarification required on the types of materials	Metals like Copper, Aluminum, Steel, Titanium and Superalloys in the range 100HBW to 700HBW is to be covered
5	5	Is table to be provided	Table dimensions have been provided as reference. Bidders don't have to provide a table. The space will be accommodated at the site