

**Corrigendum-3**  
**Tender Ref no STREAM/EQ/014/01/2025**  
**Rockwell Hardness Testing Machine**

<b>Sl no.</b>	<b>Tender clause-Annexure -B</b>	<b>Original Clause</b>	<b>Revised Clause</b>
1	13	Diamond indenter, Ball Indentors (1/16", 2.5mm)	Diamond indenter, Ball Indentors (1/16")
2	18	Certified Reference Materials for the required scales (low, medium and high) along with calibration certificate traceable to National / International	3 CRMs for each scale (A, B and C) Rockwell C scale should span between 22HRC to 70HRC