

Corrigendum-2
Tender Ref no STREAM/EQ/015/01/2025
Micro Vickers and Knoop Hardness Testing Machine

Sl no.	Tender clause-Annexure -B	Original Clause	Revised Clause
1	1.3	The hardness equipment should be capable of a range of test loads to enable performing the micro-Vickers hardness, Knoop hardness as well as macro-Vickers hardness, in the range of 10 gf to 10 kgf.	“bidder shall provide measures to ensure that good indentation at 10gf can be taken with no distortions”
2	1.13	An 18-megapixel camera with a 4K full-colour integrated system controller with an Intel (or equivalent) i7 processor, 1 Tb HDD or 512 Gb SSD (RAID configuration).	“A 10-megapixel camera minimum with a 4K full-colour integrated system controller with an Intel (or equivalent) i7 processor, 1 Tb HDD or 512 Gb SSD (RAID configuration).”
3	1.10	Manual XY stage with analogue metric micrometers and essential connector plates required (standard accessory). Stage should be 100 x 100 mm with 25 x 25 mm displacement, analog micron meter least count of 0.01mm.	“ Computerized XY stage with a least count of 0.01mm

