

**Prebid online meeting questions and reply**  
**Tender Ref no STREAM/EQ/015/01/2025**  
**Micro Vickers and Knoop Hardness Testing Machine**

<b>Sl no.</b>	<b>Tender clause-Annexure -B</b>	<b>Question</b>	<b>Reply</b>
1	1.7	For motorized turret will the total number of positions be 4?	Yes, minimum of 4 positions is required up to 6
2	1.3	Will vibration table be required since low loads of 10gf will be used	Ref. corrigendum 2
3	1.13	Is 18MP camera required if the sample stage is not going to be automatic	Refer corrigendum 2