

Prebid online meeting questions and reply
Tender Ref no STREAM/EQ/016/01/2025
UPRIGHT METALLURGICAL MICROSCOPE

Sl no.	Tender clause-Annexure -B	Question	Reply
1	3.05	LED light is supplied for which filter is not required. It will be applicable for halogen light.	Refer Corrigendum 4
2	3.06	The viewing angle is restricted and each manufacturer will have minor differences	Refer Corrigendum 4
3	3.08	C-DIC up gradation may not be required and limits the range of microscopes	Refer corrigendum 4
4	3.09	Semi-apochromatic lens will provide better clarity	Refer corrigendum 4
5	4.00	Re-look into the pixel size	Refer corrigendum 4
6	Page 23 of 25, software line item 2	Number of phases is very high when in application fewer phases will be used	Refer corrigendum 4