

**Corrigendum - 5**  
**Tender Ref no STREAM/EQ/019/01/2025**  
**FT-Infrared Spectrometer with ATR and Pallet Press**  
**Date 21/03/2025**

<b>Sl no.</b>	<b>Tender clause-Annexure -B</b>	<b>Original Clause</b>	<b>Revised Clause</b>
1	3	Helium-Neon with 10 year warranty	Helium neon or solid state laser with 10 years warranty
2	4	Michelson's Interferometer with dynamic alignment (preferred 30° angle).	– Michelson's Interferometer with dynamic or permanently aligned
3	9	0.3 cm-1 to 16 cm-1 or better, with minimum 7 or more steps.	– 0.75 cm-1 to 16 cm-1 or better with in minimum 7 or more steps
4	17	Solid, liquid, powder, coated material, nano material, paste, thin film, etc.	Solid, liquid, powder, coated material, thin film, etc.
5	21	Includes hydraulic press, mortar, pestle, etc., for KBr pellet preparation (manual).	Point Removed.

6	25	5 kVA Online UPS with 30 minutes backup.	5 kVA Online UPS or suitable UPS with 30 minutes backup.
7	29	Upgradable with single and Raman microscope instruments for research-based applications in the future.	Point Removed.