

STREAM FOUNDATION

Tender Ref. No: STREAM/EQ/061/12/2025

PARTICLE SIZE ANALYSER

CORRIGENDUM-1

DATE-28/01/2026

Sl No	Tender clause – ANNEXURE B	Original clause	Revised clause
1	Point No. 3 Optical Module .C	All lasers must be solid state diode lasers, for better stability and coherence, LED light sources not recommended.	All lasers must be solid state diode lasers/gas lasers, for better stability and coherence, LED light sources not recommended.
2	Point No. 3 Optical Module. D	Should have 150 or more detector elements to collect scattered light from a wide angle of 0.02° to 165° , for better sensitivity and better resolving power	Should have 60 or more detector elements to collect scattered light from a wide angle of 0.02° to 160° or better, for better sensitivity and better resolving power.