

STREAM FOUNDATION

Tender Ref. No: STREAM/EQ/057/12/2025

Coordinate Measuring System

CORRIGENDUM-1

DATE-27/01/2026

SI No	Tender clause – ANNEXURE B	Original clause	Revised clause
1	3.1	Probe Head Spec (B1–B10) Single & multipoint scanning, 5° indexing, A: +90°, B: ±180°, Stylus L: 300 mm, W: 33g+, Ø min: 0.5 mm, auto qualifying, optical/laser sensor compatibility.	Probe Head Spec (B1–B10) Single & multipoint scanning, 5° indexing, A: +90°, B: ±180°, Stylus L: 150 mm or above, W: 33g+, Ø min: 0.5 mm, auto qualifying, optical/laser sensor compatibility.
2	4.1	Length Measurement Error 1.9 + L/250 µm (Indexable head)	Length Measurement Error 2.3+3.3L/1000 µm
3	4.2	Repeatability 2.0 µm	Repeatability 2.2 µm
4	4.3	Single Stylus Form Error 1.9 µm	Single Stylus Form Error 2.3 µm
5	4.4	Scanning Probing Error 3.3 µm in 49 sec	Scanning Probing Error 4.5 µm in 120 Sec
6	5.3	Drive Motorized, joystick, speed 450 mm/s+, accel 1600 mm/s ² +	Drive Speed 500mm/s Acceleration: 1500 mm/s ²